



DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

m5027n

PUNGED RYK

Food and Drug Administration
Minneapolis District
240 Hennepin Avenue
Minneapolis MN 55401-1999
Telephone: 612-334-4100

January 5, 2001

WARNING LETTER

CERTIFIED MAIL
RETURN RECEIPT REQUESTED

Refer to MIN 01 - 23

Terry Finzen
Hospital Administrator
Regions Hospital
640 Jackson Street
St. Paul, Minnesota 55101

Dear Mr. Finzen:

On December 27, 2000, a representative of the State of Minnesota, acting on behalf of the Food and Drug Administration (FDA), inspected your facility (FDA certificate #140244). This inspection revealed a serious regulatory problem involving the mammography at your facility.

Under a United States Federal law, the Mammography Quality Standards act of 1992 (MQSA), your facility must meet specific requirements for mammography. These requirements help protect the health of women by assuring that a facility can perform quality mammography. Based on the documentation your site presented at the time of the inspection, the following Level 1 and Level 2 findings were documented at your facility;



Level 1 Non-Compliances:

1. Processor QC records were missing for nine consecutive days for the mammography film processor (Kodak, RPX-OMAT M6B, 6AN, 6AW, Room #1, Main).
2. Processor QC records were missing for nine out of 22 days of operation (41%) during June 2000 (~~~~~ Room #1, Main).

Page Two

Terry Finzen
January 5, 2001

Level 2 Non-Compliances:

3. The written procedure for handling mammography consumer complaints is incomplete in that it lacks all of the mandatory elements.
4. Corrective action before further exams, for a failing image score, or a phantom background optical density, or density difference outside the allowable regulatory limits, was not documented for unit 1   Mammography Room #1261).

The specific problems noted above appeared on your MQSA Facility Inspection Report which was issued to your facility following the close of the inspection.

Because these conditions may be symptomatic of serious underlying problems that could compromise the quality of mammography at your facility, they represent a serious violation of the law which may result in FDA taking regulatory action without further notice to you. These actions include, but are not limited to, placing your facility under a Directed Plan of Correction, charging your facility for the cost of on-site monitoring, assessing civil money penalties up to \$10,000 for each failure to substantially comply with, or each day of failure to substantially comply with, the Standards, suspension or revocation of your facility's FDA certificate, or obtaining a court injunction against further mammography.

It is necessary for you to act on this matter immediately. Please explain to this office in writing within 15 working days from the date you received this letter:

- the specific steps you have taken to correct all of the violations noted in this letter;
- each step your facility is taking to prevent the recurrence of similar violations;
- equipment settings (including technique factors), raw test data, and calculated final results, where appropriate; and
- sample records that demonstrate proper record keeping procedures if the findings relate to quality control or other records.

Please submit your response to Thomas W. Garvin, Radiological Health Specialist, Food and Drug Administration, 2675 No. Mayfair Road, Suite 200, Milwaukee, WI 53226-1305.

Finally, you should understand that there are many FDA requirements pertaining to mammography. This letter pertains only to findings of your inspection and does not necessarily address other obligations you have under the law. You may obtain general information about all of FDA's requirements for mammography facilities by contacting the Mammography Quality Assurance Program, Food and Drug


Page Three

Terry Finzen
January 5, 2001

Administration, P.O. Box 6057, Columbia MD 21045-6057 (1-800-838-7715) or
through the Internet at <http://www.fda.gov/cdrh/mammography/index.html>.


If you have specific questions about mammography facility requirements or about
the content of this letter please feel free to phone Mr. Garvin at (414) 771-7167
ext. 12.

Sincerely,



David R. Yost
Acting Director
Minneapolis District

TWG/ccl

xc: 
Lead Interpreting Radiologist
Regions Hospital
640 Jackson Street
St. Paul, MN 55101

Sue McClanahan
Supervisor, Radiation Unit
Minnesota Department of Health
1645 Energy Park Drive, Suite 300
St. Paul, MN 55108-2970

Priscilla F. Butler
Director, Breast Imaging Accreditation Programs
American College of Radiology
1891 Preston White Drive
Reston, VA 20191